

BGM3 zero-damp test report

Lanagan, vessel: LSSL

Release Date: 2016/07/18 17:12:00 UTC

Sensor: S220

Software version: 1.2

Port/Pier/Berth: BIO - South

Axis	pitch
Periods	4
Test start	2016/07/18 17:02:48.083
T1	2016/07/18 17:05:00.290
T2	2016/07/18 17:07:16.095
T3	2016/07/18 17:09:26.240
Test end	2016/07/18 17:11:36.395
Tolerance	5 s
4T Error	-4.688 s

Table 1: Timings

Action: Performance within specification and tolerance. No action required.