

BGM3 zero-damp test report

Lanagan, Faluotico, vessel: LSSL

Release Date: 2016/11/18 18:22:23 UTC

Sensor: S220

Software version: 1.2

Port/Pier/Berth: BIO - SOUTH

Axis	pitch
Periods	4
Test start	2016/11/18 18:13:30.557
T1	2016/11/18 18:15:45.579
T2	2016/11/18 18:17:57.856
T3	2016/11/18 18:20:07.463
Test end	2016/11/18 18:22:20.980
Tolerance	5 s
4T Error	-2.577 s

Table 1: Timings

Action: Performance within specification and tolerance. No action required.