

BGM3 zero-damp test report

Lanagan, vessel: LSSL

Release Date: 2016/07/20 16:29:53 UTC

Sensor: S220

Software version: 1.2

Port/Pier/Berth: BIO - South

Axis	roll
Periods	4
Test start	2016/07/20 16:20:57.389
T1	2016/07/20 16:23:10.619
T2	2016/07/20 16:25:26.645
T3	2016/07/20 16:27:37.833
Test end	2016/07/20 16:29:50.491
Tolerance	5 s
4T Error	0.102 s

Table 1: Timings

Action: Performance within specification and tolerance. No action required.