

BGM3 tilt test report

Lanagan, Faluotico, Oakey, vessel: LSSL

Release Date: 2016/11/18 17:41:39 UTC

Sensor: S220

Software version: 1.2

Port/Pier/Berth: BIO South

Axis	pitch
Filtered gravity 1	980563.05
Filtered gravity 2	980563.04
Misalignment	0.0037777033 mrad
Tolerance	0.7mrad

Table 1: Gravity readings and misalignment

Action: Performance within specification and tolerance. No action required.