

# BGM3 tilt test report

Lanagan, Faluotico, vessel: LSSL

Release Date: 2016/11/18 18:10:07 UTC

Sensor: S220

Software version: 1.2

Port/Pier/Berth: BIO South

Axis	pitch
Filtered gravity 1	980561.76
Filtered gravity 2	980562.63
Misalignment	-0.30995157 mrad
Tolerance	0.7mrad

Table 1: Gravity readings and misalignment

Action: Performance within specification and tolerance. No action required.