

BGM3 tilt test report

Lanagan, Faluotico, Oakey, vessel: LSSL

Release Date: 2016/11/18 17:59:13 UTC

Sensor: S220

Software version: 1.2

Port/Pier/Berth: BIO South

Axis	roll
Filtered gravity 1	980562.14
Filtered gravity 2	980562.24
Misalignment	-0.03399933 mrad
Tolerance	0.7mrad

Table 1: Gravity readings and misalignment

Action: Performance within specification and tolerance. No action required.