



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Woods Hole Oceanographic Institution
Job Number 63240
Date 3/8/2011
Technician DG

<u>Serial Number</u>	4530841-0063
----------------------	--------------

<u>Low Pressure (PSI)</u>	30	PSI
---------------------------	----	-----

<u>Time (Minutes)</u>	30	Minutes
-----------------------	----	---------

<u>High Pressure (PSI)</u>	*N/A	PSI
----------------------------	------	-----

<u>Time (Minutes)</u>	*N/A	Minutes
-----------------------	------	---------

Pass

Fail

Comments

Replaced the main piston "O"-Rings.
 *The unit is not designed for high pressure applications.

